

agilent



1050, 1100, 1200

optimax check valves

Optimize P/N	Optimize Description	Agilent P/N
27-56-01330	OPTI-MAX Outlet Cartridge, CV 1/8" Ceramic Ball & Seat, SS Cartridge	G1312-60012 G1311-60012
27-58-01330	OPTI-MAX Outlet Cartridge CV, 1/8" Ceramic Ball & Seat, PEEK Cartridge	G1312-60012 G1311-60012
27-38-02363	Replacement Cartridge for Active Inlet Check Valve	5062-8562
27-34-00658	Sapphire Piston	5063-6586
27-36-01570	OPTI-SEAL UHMW-PE Piston Seal	0905-1420
27-36-01571	OPTI-SEAL UHMW-PE Piston Seal, 10pk	n/a
27-36-01572	ITB PTFE Piston Seal	5063-6589
27-36-01573	ITB PTFE Piston Seal, 10pk	n/a
27-36-02364	OPTI-SEAL UHMW-PE Wash Seal	0905-1175
27-36-02365	OPTI-SEAL UHMW-PE Wash Seal, 10pk	n/a
27-36-02366	ITB PTFE Wash Seal	0905-1175
27-36-02367	ITB PTFE Wash Seal, 10pk	n/a
27-36-02851	Gasket Wash Seal for 1050, 1100, 1200, 6pk	5062-2484

purge valve maintenance

Optimize P/N	Optimize Description	Agilent P/N	
27-31-02277	Purge Valve Upgrade Kit for 1050, 1100	n/a	
27-31-02316	Replacement PEEK Plug for HP/Agilent PTFE Filter	n/a	
27-31-02853	2-in-1 Seal Cap for Agilent 1050, 1100, 1200	Replaces Agilent Gold Seal and Outlet Cap	5001-3707 5062-2485
27-04-02852	PTFE Frits for 1050, 1100, 1200 6/pk	01018-22707	

autosampler

Optimize P/N	Optimize Description	Agilent P/N
27-31-03212	Rotor Seal Assembly for 1100, 1200 (Vespel)	0100-1853
27-31-03215	Needle, 1100, 1200	G1313-87201
27-31-03329	Needle, Seat Assembly, 1100, 1200	G1313-87101
27-31-03603	Vespel Rotor Seal, 1050	0101-0626
27-31-03146	Needle Seat Assembly 2 pc Ferrule	

lamps

Optimize P/N	Optimize Description	HITACHI P/N
27-32-02817	Deuterium Lamp, 1050 VWD A-Series, 79853A	79853-60002
27-32-02827	Deuterium Lamp, 1100 CE DAD	2140-0585
27-32-02844	Deuterium Lamp, Agilent/HP 1050C, 1050DAD, 1090, 1040	79883-60002



optimize technologies

13993 Fir Street
Oregon City, OR 97015

TF: 800.669.9015 | Tel: 503.557.9994 | Fax: 503.557.9995

www.optimizeotech.com